

IR319B Single and IR320B Dual Patient Stations

IL713 Specification Sheet

Tek-CARE®300II Nurse Call System • Rev. 5 - 11/2016

Features

- UL® 1069 Listed
- Supervised
- All stations interchangeable
- Simplified multiplex wiring and plug-in connectors
- Input available for Hill-Rom® Sidecom® beds or remote call-in button applications
- Uses pillow speakers or push-button call cords

Description

The IR319B Single Patient Station provides reliable two-way, hands-free communication between patient and nurse, and provides for programmable privacy, paging and patient priority from the master station. Calling the nurse is accomplished by pressing the call cord button or the area indicated nurse on the pillow speaker. Solid state LEDs on the panel indicate the call has been placed and the intercom is in use.

If the pillow speaker or call cord is accidentally removed from its receptacle, the nurse master station registers a non-cancellable cord out call. The call may be reset only by returning the call cord or pillow speaker to its receptacle.

The IR320B Dual Patient Station has the same specifications and functions as the IR319B, plus an additional independently operating call cord/pillow speaker receptacle and a second call-placed indicator.

Listings



UL® 1069 listed; UL® File Number E71599.



IR319B Single and IR320B Dual Patient Stations

Specifications

Dimensions:	Height:	4.5"	(114 mm)
	Width:	6.375"	(162 mm)
	Depth:	2.75"	(70 mm)
	Projects:	1.0"	(25 mm)
	above mounting surface		
Construction:	Flame-retardant (UL® 94V-0) plastic faceplate		
Speaker:	2.5" × 2.5" (64 mm × 64 mm)		
Connections:	Plug-in type		
Housing:	Steel City #H3BD and #3GC plaster ring, or exact equal. Horizontal mount.		

Required Components

Tek-CARE®300II Nurse Call System